

## CORRIGENDUM

**(Tender Notice No. DCQC/50/2015, MUMBAI-65 and E-Tender Notice no. 693 Dt.21.01.2015)**

Vide Tender Notice No. DCQC/QCL/APEDA/INSTRUMENT/50/2015, MUMBAI-65 and **E-Tender Notice no. 693 Dt.21.01.2015**, tender for purchase of **four** Laboratory instruments for the office of Deputy Commissioner Animal Husbandry, Quality Control Laboratory, Aarey , Goregaon, Mumbai-65, was floated on 21.01.2015. According to the discussion in the pre-bid meeting dated 29.01.2015, Commissioner Animal Husbandry, M.S., Pune-7 has proposed to consider the following changes in **Annexure A- Technical Specifications**.

<b>Sr. No</b>	<b>Name of the instrument</b>	<b>Point no. and existing lines of Specification</b>	<b>Amended Text</b>
1(B)	GC-MS/MS System	<b>1.ION SOURCE (line no. 3)</b> The inner source housing must be isolated to allow rapid source maintenance <b>without the need to vent the system.</b>	<b>1.ION SOURCE (line no. 3)</b> The inner source housing must be isolated to allow rapid source maintenance.
1(c)	Atomic Absorption Spectrophotometer	<b>1(C).1.2</b> Diffraction grating with <b>1800</b> lines used as single/dual blazed with sensitivity of 0.9 absorbance for 5 ppm copper solution and high monochromator efficiency	<b>1(C).1.2</b> Diffraction grating with <b>suitable</b> lines used as single/dual blazed with sensitivity of 0.9 absorbance for 5 ppm copper solution and high monochromator efficiency

Except above points, all other points remain unchanged in **Tender notice no. 50/2015 and E-Tender notice no. 693 Dt.21.01.2015**. The Bidders are required to take a note of the above mentioned amendments.

This Corrigendum and amendments shall be considered a part of the Original Tender Document. Note be taken of this.

**Commissioner**  
**Department of Animal Husbandry,**  
**Maharashtra State, Pune 411 007**