

CORRIGENDUM

Sub: Supply of Precision/Analytical weighing balance (3 Nos.) under RKVY funded project for the year 2017-2018

Ref: **E- Tender No. 1624**, Tender Notice No. BVC/RKVY/TVCC/TENDER/124/2017

No.	Details	Existing	Revised
1]	Technical Specifications on Page No. 11 and Clause by Clause on Page No. 22	Readability 0.0001 mg and Linearity ± 0.0002 mg	Readability 0.0001 gm and Linearity ± 0.0002 gm.
2]	Technical Specifications on Page No. 11 and Clause by Clause on Page No. 22	Internal Calibration	Fully automatic calibration with temperature and time control along with interface port

Sd/-

Associate Dean
Bombay Veterinary College